

8 March 2016

WM Project Number: 06159-WM Our Ref: AGL080316RH Email: aclifton@agl.com.au

Aaron Clifton AGL Upstream Investments Pty Ltd Lot 35, Medhurst Road MENANGLE NSW 2568

Dear Aaron

### Re: Operational Noise Monitoring of Well Sites SF17 and SF20

#### Introduction

Wilkinson Murray was commissioned to conduct operational noise monitoring of well sites Spring Farm 17 (SF17) and Spring Farm 20 (SF20). Measurements were conducted at the potentially most affected residential receivers in different directions to assess compliance. This letter report summarises the results of the measurements conducted during the day, evening and night time on Wednesday, 2 March 2016.

Both well sites were free flowing during the measurements. At SF17, all three (3) wells (SF01, SF02 and SF03) were producing. At SF20, three wells (SF05, SF07 and SF08) were producing. During the day (before 6.00pm), a workover rig and ancillary equipment were also operating at well SF09.

It is important to note that civil works associated with the construction of Liz Kernohan Drive and Spring Farm residential estate to the east of Richardson Road was taking place during daytime hours (before 6.00pm).

#### **Measurement Locations**

Measurements conducted at the potentially most affected residential receivers surrounding the well sites are described as follows:


- R1 Corner of Holland Drive and Reynolds Street (representing the easternmost point of Landcom's Spring Farm Eastern Village residential development area currently in construction)
- R2 67 Dewpoint Drive (existing house located directly west of the well sites)
- R3 15 Jane Court (existing house located to the north of the well sites)

Figure 1 shows the measurement locations and both well sites.

Wilkinson Murray Pty Limited · ABN 39 139 833 060

Level 4, 272 Pacific Highway, Crows Nest NSW 2065, Australia • Offices in Orange, Qld & Hong Kong t +61 2 9437 4611 • f +61 2 9437 4393 • e acoustics@wilkinsonmurray.com.au • w www.wilkinsonmurray.com.au

# ACOUSTICS AND AIR



### Figure 1 Measurement Locations & Well Sites

## **Operational Noise Criteria**

The Minister's Conditions of Approval (PA06\_0291) for the Project state Project-specific operational noise criteria as shown in Table 1.

### Table 1 Project-Specific Operational Noise Criteria

		L <sub>Aeq,15min</sub> (dBA)	
Well Site	Day	Evening	Night
	(7.00am – 6.00pm)	(6.00pm – 10.00pm)	(10.00pm – 7.00am)
SF17 & SF20	43	41	36

#### **Monitoring Procedure**

Noise monitoring was conducted during the three (3) time periods defined in Table 1 in order to assess compliance for the day, evening and night time periods.

Daytime, evening and night measurements were carried out on Wednesday, 2 March 2016.

The measurements were made using a B&K Type 2236 Sound Level Meter. The sound level meter holds a current NATA calibration certificate (Calibration Certificate valid for 2 years is attached to this letter).

The reported measurements were conducted in suitable meteorological conditions (wind speeds below 5m/s and no rain). Wind speed and direction was determined using a hand-held digital anemometer Digitech QM1642. Cloud cover was observed to be 0-5%.

#### **Results of Attended Noise Measurements**

Table 2 summarises the measurement results.

#### Table 2 Measurement Results at Residential Receivers

Period	Measurement	Location	Comments	Estimated L <sub>Aeq,15min</sub> due	L <sub>Aeq,15m</sub>	<sub>nin</sub> Noise ( (dBA)	Criteria
Fenou	Start Time	Location	comments	to Well Sites (dBA)	Day	Eve	Night
Day	3.44pm	R1	Distant traffic hum audible most of the time, 40-42dBA. Civil works audible most of the time, 40-42dBA. SF20 workover generator audible most of the time, 41-43dBA. Metal clanging noise and hand tools associated with workover activities audible very briefly, 46-50dBA. SF17 well site inaudible at all times.	42dBA	43	-	-
	4.08pm	R2	Measurement dominated by civil works, 40-54dBA. Well sites inaudible at all times.	<30dBA	43	-	-
	4.38pm	R3	Measurement dominated by distant traffic, 38-44dBA, and insect noise, 35-40dBA. Well sites inaudible at all times.	<30dBA	43	-	-
Evening	6.53pm	R2	Measurement dominated by distant traffic, 37-42dBA, and bird noise, 45-50dBA. Well sites inaudible at all times.	<30dBA	-	41	-

Deviad	Measurement		<b>2</b>	Estimated L <sub>Aeq,15min</sub> due	L <sub>Aeq,15m</sub>	nin Noise ( (dBA)	Criteria
Period	Start Time	Location	Comments	to Well Sites (dBA)	Day	Eve	Night
	7.19pm	R3	Measurement dominated by distant traffic, 40-41dBA, and bird noise, 40-43dBA. Well sites inaudible at all times.	<30dBA	-	41	-
	7.46pm	R1	Measurement dominated by distant traffic, 37-38dBA. Well sites inaudible at all times.	<30dBA	-	41	-
	10.00pm	R2	Measurement dominated by cicadas, 44-47dBA. Well sites inaudible at all times.	<34dBA	-	-	36
Night	10.23pm	R3	Measurement dominated by distant traffic, 34-37dBA, and insect noise, 36-38dBA. Well sites inaudible at all times.	<30dBA	-	-	36
	10.50pm	R1	Measurement dominated by cicadas, 40-53dBA. Well sites inaudible at all times.	<30dBA	-	-	36

#### Conclusion

On the basis of the attended measurements conducted at the potentially most affected residential receivers in different directions, noise levels generated by well sites SF17 and SF20 are found to comply with the relevant operational noise criteria during the day, evening and night time periods.

I trust this information is sufficient. Please contact us if you have any further queries.

Yours faithfully WILKINSON MURRAY

RH

Roman Haverkamp Senior Engineer

#### Note

All materials specified by Wilkinson Murray Pty Limited have been selected solely on the basis of acoustic performance. Any other properties of these materials, such as fire rating, chemical properties etc. should be checked with the suppliers or other specialised bodies for fitness for a given purpose. The information contained in this document produced by Wilkinson Murray is solely for the use of the client identified on the front page of this report. Our client becomes the owner of this document upon full payment of our **Tax Invoice** for its provision. This document must not be used for any purposes other than those of the document's owner. Wilkinson Murray undertakes no duty to or accepts any responsibility to any third party who may rely upon this document. **Quality Assurance** 

We are committed to and have implemented AS/NZS ISO 9001:2008 "Quality Management Systems – Requirements". This management system has been externally certified and Licence No. QEC 13457 has been issued.

#### AAAC

This firm is a member firm of the Association of Australian Acoustical Consultants and the work here reported has been carried out in accordance with the terms of that membership.

Accredited Laboratory Number 1301         EXECUTION CONCUTIONES:         Marine	B B C C C C C C C C C C C C C	Head Office: Suite 2, Phone: +61
CALIBRATION CERTIFICATE         Yerd       Yerd         CUSTOMER:       Wilkinson Murray Pty Ltd         L23 Willoughby Road       Crows Nest NSW 2065         Manufacturer:       Brüel & Kjær         Sound Level Meter:       2236         Serial Number:       2173783         Microphone:       4188         Serial Number:       2173783         Microphone:       12 hours at 23 °C         Environment       Air temperature:       23.8 °C       ± 3°C         conditioning:       12 hours at 23 °C       ± 3.0 kPa         Relative Humidity:       34.7 % KH       ± 25 % KH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2         PROCEDURE:       The measurements have been performed with the assistance of:         Britel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage fator providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with EA-402 from deer of enginating from the standard, calibration method, effect of environmental conditions and any short time containtowerd.         Accredited Laboratory No: 130 1		
Certificate No: CAU1400292       Page 1         CUSTOMER:       Wilkinson Murray Pty Ltd         123 Willoughby Road       Crows Nest NSW 2065         Manufacturer:       Brüel & Kjær         Sound Level Meter:       2236         Serial Number:       2173783         Microphone:       4188         Serial Number:       2157590         CALIBRATION CONDITIONS:       Preconditioning:         Preconditioning:       12 hours at 23 °C         Environment       Air pressure:         Online Kaire       101.8 kPa         Relative Humidity:       34.7 % RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements at specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Brücl & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage fator providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordances with EA-402 from tep or providing a level of confidence of approximately 95%. The uncertainty evaluation and any short time contribution from the device under of approximately 95%. The uncertainty evaluation ane standard uncertainty multiplied by a coverage fator p		
CUSTOMER:       Wilkinson Murray Pty Ltd         123 Willoughby Road       Crows Nest NSW 2065         Manufacturer:       Brüel & Kjær         Sound Level Meter:       2236         Serial Number:       2173783         Microphone:       4188         Serial Number:       2173783         Microphone:       4188         Serial Number:       2173783         Microphone:       4188         Serial Number:       2173783         CALIBRATION CONDITIONS:       Preconditioning:         Preconditioning:       12 hours at 23 °C         conditions       Air pressure:       101.8 kPa       ± 3.0 kPa         Relative Humidity:       34.7 %RH       ± 25 %RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Brück & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty evaluation has been carride out in accordance with EA-402 from de priginiting from the standard, calibration method, effect of environmental conditions and any short time contribution from the device under contriduate a lemethod, effect of environmental conditions and any short		
123 Willoughby Road         Crows Nest NSW 2065         Manufacturer:       Brüel & Kjær         Sound Level Meter:       2236         Serial Number:       2173783         Microphone:       4188         Serial Number:       2157590         CALIBRATION CONDITIONS:         Preconditioning:       12 hours at 23 °C         Environment       Air pressure:         Air pressure:       101.8 kPa         Relative Humidity:       34.7 %RH         # 25 %RH    SPECIFICATIONS: The Sound Level Meter has been calibrated in accordance with the requirements at specified in AS 1259.1 and AS 1259.2. PROCEDURE: The measurements have been performed with the assistance of: Brücl & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009 RESULTS: The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage fastor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-402 from ele originating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution in writing of this testing authority.	Page 1 of 22	te No: CAU1400292
Sound Level Meter:       2236       Serial Number:       2173783         Microphone:       4188       Serial Number:       2157590         CALIBRATION CONDITIONS:       Preconditioning:       12 hours at 23 °C         Environment       Air temperature:       23.8 °C       ± 3°C         conditions       Air pressure:       101.8 kPa       ± 3,0 kPa         Relative Humidity:       34.7 %RH       ± 25 %RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-4/02 from elearing from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the device under contribution from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the standards, calibration method, effect of environmental conditions	load	123 Willoughby F
Sound Level Meter:       2236       Serial Number:       2173783         Microphone:       4188       Serial Number:       2157590         CALIBRATION CONDITIONS:       Preconditioning:       12 hours at 23 °C         Environment       Air temperature:       23.8 °C       ± 3°C         conditions       Air pressure:       101.8 kPa       ± 3,0 kPa         Relative Humidity:       34.7 %RH       ± 25 %RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-4/02 from elearing from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the device under contribution from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the standards, calibration method, effect of environmental conditions	ATTX	Denal 9 King
Microphone:       4188       Serial Number:       2157590         CALIBRATION CONDITIONS:       Preconditioning:       12 hours at 23 °C         Environment       Air temperature:       23.8 °C       ± 3°C         conditions       Air pressure:       101.8 kPa       ± 3.0 kPa         Relative Humidity:       34.7 %RH       ± 25 %RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Bruel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-402 from eleapriginating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution transport of this testing authority.	Serial Number: 2173783	
CALIBRATION CONDITIONS:         Preconditioning:       12 hours at 23 °C         Environment       Air temperature:       23.8 °C       ± 3°C         conditions       Air pressure:       101.8 kPa       ± 3.0 kPa         Relative Humidity:       34.7 % RH       ± 25 % RH         SPECIFICATIONS: The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE: The measurements have been performed with the assistance of:         Bruel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS: The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-402 from elea originating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under contribution from the published or reproduced except in full without the permission in writing of this testing authority.		
Preconditioning:       12 hours at 23 °C         Environment       Air temperature:       23.8 °C       ± 3°C         conditions       Air pressure:       101.8 kPa       ± 3.0 kPa         Relative Humidity:       34.7 %RH       ± 25 %RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor moviding a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-4/02 from elea priginating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under or calibration yno: 1301 accredited for compliance with ISO/IEC 17025.         This document may not be published or reproduced except in full without the permission in writing of this testing authority.		
Environment conditions       Air temperature:       23.8 °C       ± 3°C         Conditions       Air pressure:       101.8 kPa       ± 3.0 kPa         Relative Humidity:       34.7 %RH       ± 25 %RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-4/02 from eleapriginating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under or accordance y No: 1301 accredited for compliance with ISO/IEC 17025.         This document may not be published or reproduced except in full without the permission in writing of this testing authority.		
conditions       Air pressure:       101.8 kPa       ± 3.0 kPa         Relative Humidity:       34.7 %RH       ± 25 %RH         SPECIFICATIONS:       The Sound Level Meter has been calibrated in accordance with the requirements as specified in AS 1259.1 and AS 1259.2.         PROCEDURE:       The measurements have been performed with the assistance of:         Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009         RESULTS:       The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-4/02 from eleating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under or Accredited Laboratory No: 1301 accredited for compliance with ISO/IEC 17025.         This document may not be published or reproduced except in full without the permission in writing of this testing authority.	22.8.80	
specified in AS 1259.1 and AS 1259.2.  PROCEDURE: The measurements have been performed with the assistance of: Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009  RESULTS: The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-4/02 from elea originating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under o Accredited Laboratory No: 1301 accredited for compliance with ISO/IEC 17025. This document may not be published or reproduced except in full without the permission in writing of this testing authority.	101.8 kPa ± 3.0 kPa	Air pressure:
Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 usin procedure B&K proc 2236-4188-009 <b>RESULTS:</b> The reported expanded uncertainty is bidirectional and is based on the standard uncertainty multiplied by a coverage factor providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with EA-4/02 from elec originating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under of Accredited Laboratory No: 1301 accredited for compliance with ISO/IEC 17025. This document may not be published or reproduced except in full without the permission in writing of this testing authority.	leter has been calibrated in accordance with the requirements as	
This document may not be published or reproduced except in full without the permission in writing of this testing authority.	on System B&K 3630 with application software type 7763 using idirectional and is based on the standard uncertainty multiplied by a coverage factor k = 2 be uncertainty evaluation has been carried out in accordance with EA-4/02 from elements	er Sound Level Meter Calibrati &K proc 2236-4188-009 
Date of Calibration: 03/11/2014 Certificate Issued: 04/11/20		
	Certificate Issued: 04/11/2014	Calibration: 03/11/2014
Approved signatory:	Jan Rasmussen	